

EC Declaration of conformity

COMPANY: ACRTIKO A/S **PRODUCT INFORMATION: BIOMEDICAL FREEZERS**

ADRESS:

ODDESUNDVEJ 39, 6715 ESBJERG, DENMARK **MODEL:** LAF 700

WE HEREBY DECLARE THAT THE BELOW MENTIONED PRODUCTS COMPLY WITH THE FOLLOWING DIRECTIVES, NATIONAL LAWS AND HARMONIZED STANDARDS:

DIRECTIVES AND NATIONAL LAWS:

2006/42/EC **MACHINERY DIRECTIVE** 2011/65/EU **ROHS DIRECTIVE**

2014/30/EU ELECTROMAGNETIC COMPATIBILITY (EMC) DIRECTIVE

2014/35/EU LOW VOLTAGE DIRECTIVE

HARMONIZED STANDARDS:

EN ISO 9001:2015 QUALITY MANAGEMENT SYSTEM

EN 13485:2016 MEDICAL DEVICES QUALITY MANAGEMENT SYSTEM EN 60335-2-89:2010 / A1:2016 / A1:2017 HOUSEHOLD AND SIMILAR ELECTRICAL APPLIANCES SAFETY REQUIREMENTS FOR ELECTRICAL EQUIPMENT EN 61010-1: 2010 FOR MEASUREMENT, CONTROL, AND LABORATORY USE

PART 1: GENERAL REQUIREMENTS

EMISSION:

EN 55014-1:2006, A1:2009, A2:2011 PRODUCT FAMILY STANDARD - REQUIREMENTS FOR

HOUSEHOLD APPLIANCES, ELECTRIC TOOLS AND

SIMILAR APPARATUS

EN 61000-3-3:2013 PRODUCT FAMILY STANDARD - FLICKER

EN 61000-3-2:2014 PRODUCT FAMILY STANDARD - HARMONIC CURRENT

EN 61326-1:2013 PRODUCT FAMILY STANDARD, MEASUREMENT,

CONTROL AND LABORATORY EQUIPMENT

IMMUNITY:

EN 61326-1:2013

PRODUCT FAMILY STANDARD REQUIREMENTS EN 55014-2:1997 + Corr.:1998, A1:2001, A2:2008

FOR HOUSEHOLD APPLIANCES, ELECTRIC TOOLS AND

SIMILAR APPARATUS - PART 2: IMMUNITY

PRODUCT FAMILY STANDARD MEASUREMENT,

CONTROL AND LABORATORY EQUIPMENT

Esbjerg 01/03/2021

Tove Tækker Svendsen

R&D & QA Manager

5082033-05